

# Removing and Replacing Air Filter

*Ventaire™ Endoscopy System*

*Instructions for Scope Drying Unit*

Follow these instructions to remove and replace the air filter on the scope drying unit.

**To remove:**

Press the quick release buttons on the top and bottom of the air filter, and pull out and discard the filter

**To replace:**

Insert the new air filter. The quick release buttons will reset once the filter is attached

