

# Removing and Replacing Tubes

## Ventaire™ Endoscopy System

### Instructions for Scope Drying Unit

---

Follow these instructions to remove and replace drying tubes.

**To remove:**

Turn the tube fitting to the left to loosen and remove the tube from the port on the scope drying unit

**To replace:**

Connect a new tube to the port by turning the tube fitting to the right until tube is firmly attached

